

Bacha Khan Medical Complex - MTI, Swabi

Technical Evaluation of CT Scan

S#	Item Name	Firm	Model	Make	Conformance Specifications												Total Marks	Remarks																
					Compliance to Purchaser's Specifications			Special Features			Product Certification			Performance Certification					Firm Evaluation															
					28	2	FDA CE JIS / MHLW			Valid ISO 9001 Quality Management Certificate Valid ISO 13485 Quality management certificate Valid ISO 45001 Quality management certificate			Satisfactory Performance Certificate for each quoted Model from medical institutions within Pakistan Satisfactory Performance Certificate for Previous quoted Model from medical institutions of Pakistan		Manufacturer Authorization Certificate (PHC / Embassy Attested) Satisfactory performance certificate(s) of the firms Experience above 5 years issued by well reputable organizations Most recent Audti report IT-1 or IT-2 Form showing net annual sales			Diploma Engineer Graduate Engineers Bio Medical Engineer / Foreign trained Engineer for the quoted equipment			Availability of workshop in Peshawar Availability of workshop at National level		Certificate to the affect that the firm will provide 95% uptime on company letter head. List of tools, testing equipment and calibration equipment relevant to the product Spare Parts readily available beyond ten years from the manufacturer		Warranty Period 3 years both with spare parts and services from the date of successful installation is mandatory. Additional warranty with parts and services for Each year (2 marks)		(Lowest / Quoted) *4 Including Parts & Services							
		Medequips	Aquilon Start	CANON MEDICAL SYSTEM		28	2	3	3	3	1	1	1	4	5	Mandatory	2	1	1	1	1	2	Mandatory	1	Mandatory	1	1	Mandatory	4	L	Q	Marks	62.5176794	
1	CT Scan Machine	Siemens	SOMATOM go Now	Siemens Healthineers		26	2	3	3	0	1	1	1	3	5	Provided	2	1	1	1	1	2	Provided	1	Provided	1	1	Provided	0	78	78	4	60	
		Shirazi	Optima CT 520	Siemens Healthineers																														

Disqualified due to Major Deviation
 Hospital Specs Detector width should be 16mm minimum
 Vender specs 0.7mmx16mm = 11mm